



# **Declaration of conformity**

F	or	the	foll	owina	eaui	pment	:

Product Name: AC/DC Medical Adaptor

Model Designation: GSM120Ax (x=12,15,20,24,48)

is herewith confirmed to comply with the requirements set out in the Council Directive, the following standards were

# RoHS Directive (2011/65/EU), (EU)2015/863

### MDR Directive (EU) 2017/745

EN 60601-1:2006+A11+A1+A12 TUV certificate No: TA 50338894

# **MDR Directive (EU) 2017/745**

EN 60601-1-2:2015

### **EMI (Electro-Magnetic Interference)**

Conducted emission / Radiated emission

	EN 55011:2016+A11:2020 EN IEC 61024-3:2018	Class B
Harmonic current	EN IEC 61000-3-2:2019+A1:2021	
Voltage flicker	EN 61000-3-3:2013+A1:2019	

EMS (Electro-Magnetic Susceptibility)							
EN 60601-1-2:2015	EN IEC 61024-3:2018						
ESD air	EN 61000-4-2:2009	Level 4	15KV				
ESD contact	EN 61000-4-2:2009	Level 4	8KV				
RF field susceptibility	EN IEC 61000-4-3:2020	Level 3	10V/m(80MHz-2.7GHz)				
RF field susceptibility	EN IEC 61000-4-3:2020	Table 9	9~28V/m (385MHz~5.78GHz)				
EFT bursts	EN 61000-4-4:2012	Level 3	2KV/100KHz				
Surge susceptibility	EN 61000-4-5:2014+A1:2017	Level 3	1KV/Line-Line				
Surge susceptibility	EN 61000-4-5:2014+A1:2017	Level 3	2KV/Line-FG				
Conducted susceptibili	ity EN 61000-4-6:2014	Level 3	10V				
Magnetic field immunit	y EN 61000-4-8:2010	Level 4	30A/m				
	EN IEC 61000-4-11:2020						
	0% recidual voltage for 0.5 avdec 0% i	ocidual valtaac	for 1 andos 70% residual voltage for 25 andos 0%				

Voltage dip, interruption

0% residual voltage for 0.5 cycles,0% residual voltage for 1 cycles ,70% residual voltage for 25 cycles, 0% residual voltage for 250 cycles

The power supply is considered as a component that will be operated in combination with final equipment. Since EMC performance will be affected by the complete system, the final equipment manufacturers must re-qualify EMC Directive on the complete system

For guidance on how to perform these EMC tests, please refer to TDF (Technical Documentation File).

The product under declaration is just a unit without medical function. Complete MDR should only be verified when it is used together with particular medical device(s).

This Declaration is effective from serial number SC1xxxxxxx

Person responsible for marking this declaration:

Mean Well Enterprises Co., Ltd.

(Manufacturer Name)

No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 24891, Taiwan

(Manufacturer Address)

Aries Jian/ Director, Group R&D:

(Name / Position)

(Signature)

Alex Tsai/Director, Product Strategy Center: (Name / Position)

(Signature)

Taiwan

Sep 14, 2021 (Date)

(Place)