



EC declaration of conformity			
For the following equipment:			
Product Name: Switching Power Supplies			
Model Designation: GSM40Ax (x=05,07,09,12,15,18,24,48), GSM60Ax (x=05,07,09,12,15,18,24,48)			
	comply with the requirements set out in the ving standards were applied:	he Council	Directive 93/42/EEC concerning
RoHS Directive (2011 Medical Directive (93/4 Low Voltage			
EN60601-1:2006	TUV certificate No : TA	A50262994	
EMI (Electro-Magnetic In Conducted emission / Rad	25 27 37 37 37 37 37 37 37 37 37 37 37 37 37		Class B Class B
Harmonic current	EN61000-3-2:2006		
Voltage flicker	EN61000-3-3: 2008		
EMS (Electro-Magnetic S	Susceptibility)		
EN60601-1-2:2007 EN5	5024:2010 EN61204-3:2000 EN61000-6	3-1:2007	
ESD air	EN61000-4-2:2009	Level 3	8KV
ESD contact	EN61000-4-2:2009	Level 2	6KV
RF field susceptibility	EN61000-4-3:2006+A1:2008+A2:2010	Level 3	10V/m
EFT bursts	EN61000-4-4:2012	Level3	2KV/5KHz
Surge susceptibility	EN61000-4-5:2006	Level 4	1KV/Line-Line 2KV/Line-Earth
Conducted susceptibility	EN61000-4-6:2009	Level 3	10V
Magnetic field immunity	EN61000-4-8:2010	Level 4	3A/m
Voltage dip, interruption	EN61000-4-11:2004 >95% dip 0.5 periods 3	80% dip 25 pe	eriods >95% interruptions 250 periods
performance will be affected complete system again.	ered as a component that will be operated in by the complete system, the final equipment ma rm these EMC tests, please refer to TDF (Techn	anufacturers	must re-qualify EMC Directive on the
This Declaration is effective for	rom serial number EB3xxxxxxx		
Person responsible for ma	rking this declaration :		
Mean Well Enterprises Co (Manufacturer Name)	., Ltd.		<u> </u>
No.28, Wuquan 3rd Rd., V (Manufacturer Address)	Vugu Dist., New Taipei City 248, Taiwan (R		MADDO
Johnny Huang/Senior Verification (Name / Position)	on Engineer : Ted Che (Signature) (Name / F	ng/Product Position)	Manager: Head Cherry (Signature)

Apr. 19, 2013 (Date)

Taiwan

(Place)