



## EC-Conformity Declaration

For the following equipment :

Switching Power Supplies

(Product Name)

MPS-y-z (y=45,65) (z=3,3,5,7,5,12,13,5,15,24,27,48)

(Model Designation)

is herewith confirmed to comply with the requirements set out in the Council Directive, the following standards were applied :

Directive 93/42/EEC concerning medical devices :

EN60601-1: MPx-yz series TUV certificate No:TA5000971301

Conducted emissions	EN55011 (EN55022)	Class B	
Radiated emissions	EN55011 (EN55022)	Class B	
ESD air	IEC801-2(EN61000-4-2)	Level 3	8KV
ESD contact	IEC801-2(EN61000-4-2)	Level 2	4KV
RF field susceptibility	IEC801-3(EN61000-4-3)	Level 2	3V/m
EFT bursts	IEC801-4(EN61000-4-4)	Level 3	2KV/5KHz
Surge susceptibility	IEC801-5(EN61000-4-5)	Level 3	1KV/Line-Line 2KV/Line-Earth
Harmonic distortion	EN61000-3-2	Conforms	
Voltage flicker	EN61000-3-3	Conforms	
Conducted susceptibility	EN61000-4-6	Level 2	3V
Magnetic field immunity	EN61000-4-8	Level 2	3A/m
Voltage dip, interruption	EN61000-4-11(EN55024)		>95% dip 0.5 periods 30% dip 25 periods >95% interruptions 250 periods
Keyed carrier immunity	ENV50204	Level 2	3V/m 900MHz

Component power supply will be installed into a final equipment. Since EMC performance will be effected by complete installation, the final equipment manufacturers must re-conform EMC Directive on final equipment again.

Importer in the EC :

Mean Well Europe B.V. Co., Ltd.

(Importer Name)

Spinnerij 73-75, 1185 ZS Amstelveen, the Netherlands

(Importer Address)

Manufacturer responsible for marking this declaration :

Mean Well Enterprises Co., Ltd.

(Manufacturer Name)

28, Wu-Chuan 3rd Road, Wu Ku Ind. Park, Taipei Hsien, Taiwan R.O.C.

(Manufacturer Address)

Roger Lai.

(Name, Surname)

Manager

(Position/Title)

Taiwan

(Place)

January.23.2009

(Date)

(Legal Signature)